

【訂正】英文版 医薬品添加物規格 2018

2019 年 8 月に発行いたしました『英文版 医薬品添加物規格 2018』の記述に誤りがございました。お詫びを申し上げますとともに、下記の通り訂正させていただきます。（2022 年 1 月 31 日現在）

該当箇所	正	誤
p.243 下から 3 行目	<u>0.70</u> mL of 0.01 mol/L hydrochloric acid VS	<u>0.07</u> mL of 0.01 mol/L hydrochloric acid VS
p.250 下から 5 行目※	Absorbance	Optical rotation
p.468 上から 12 行目※	for ferric <u>salt</u> and for chloride	for ferric <u>chloride</u> and for chloride
p.589 上から 2-3 行目	(2) Arsenic—Prepare the test solution with 2.0 g of Methacrylic Acid Copolymer LD according to <u>Method 3</u> , and perform the test (not more than 1 ppm).	(2) Arsenic—Prepare the test solution with 2.0 g of Methacrylic Acid Copolymer LD according to <u>Method 2</u> , and perform the test (not more than 1 ppm).
p.785 下から 9 行目※	and shake for about 15 <u>seconds</u> ,	and shake for about 15 <u>minutes</u> ,
p.786 下から 15 行目	Viscosity 350 — 550 <u>mm²/s</u> (Method 1, 20°C).	Viscosity 350 — 550 <u>mm²s</u> (Method 1, 20°C).
p.836 上から 10 行目※	To <u>20</u> g of Pregelatinized Starch	To <u>2.0</u> g of Pregelatinized Starch
p.858 上から 12 行目	of which value is not <u>less</u> than 1.8% and not <u>more</u> than 2.2%.	of which value is not <u>more</u> than 1.8% and not <u>less</u> than 2.2%.

※：第 1 刷の正誤。第 2 刷以降は修正済み。